


High Voltage Pulse Generator for TOF

Output Voltage(drawing-in)	-3kVp
Output Voltage(pushing-out)	300Vp
Rise Time	less than 100ns
Pulse Width	100-200 μ s
Repetition	1kHz
Load Capacitance	less than 300pF



CH1(Y):Ext Trigger	10V/div
CH3(R):Pushing-out Volatge	100V/div
CH4(B):Drawing-in Voltage	1kV/div 50 μ s/div



Pulsed Power Japan Laboratory Ltd.

Techno Factory Shiga Bldg. No. 2, 7-3-46, Noji-higashi, Kusatsu, Sjiga, 525-0058, Japan

TEL: +81-77-598-1470 FAX: +81-77-598-1490

HP: <http://www.mypj.com> Mail: info@myppj.com